


- NOTES:
- ALL CABLING TO BE RUN IN 1"-2" CONDUIT, DEPENDING ON ACTUAL WIRE SIZE USED
 - SCALE CONTROLLER REQUIRES 100-240VAC, 1PH., 50/60 HZ, 15 AMP MAXIMUM, BRANCH RATED POWER SOURCE
 - ALL ELECTRICAL WIRING AND ROUTING MUST CONFORM TO LOCAL NEC CODE
 - WHEEL DETECTOR CABLES ARE 60' LONG. IF DISTANCE FROM W/D TO CONTROLLER IS GREATER THAN 60', A JUNCTION BOX (64069679) AND ADDITIONAL CABLE (68003109, BY FOOT) ARE NEEDED.



CIM SITE LAYOUT (TOP VIEW)

METTLER TOLEDO <small>THIS PRINT IS FURNISHED WITH UNDERSTANDING THAT THE ESSENCE THEREOF WILL NOT BE REPRODUCED IN WHOLE OR IN PART WITHOUT THE WRITTEN PERMISSION OF METTLER TOLEDO. ALL DIMENSIONS ARE THE PROPERTY OF METTLER-TOLEDO, INC. AND WILL BE PROTECTED BY PATENTS.</small>		M.T. CODE FINISH CODE DIMENSION TOLERANCE UNLESS OTHERWISE SPECIFIED DECIMAL .XXX = ±.005 FRACTIONAL ±1/32 ANGULAR ±1/2° SCALE NONE DO NOT SCALE DRAWG. SUPPLIES BY SUPPLIED BY	NAME TWO PLATFORM CIM SITE LAYOUT - SAMPLE DRW: ERW CHKD: APPD 02-MAR-17 02-MAR-17 SHEET 1 OF 2
AUTOMATIC SPEC: AS SHOWN FINISH DESCRIPTION DIMENSION TOLERANCE UNLESS OTHERWISE SPECIFIED DECIMAL .XXX = ±.005 FRACTIONAL ±1/32 ANGULAR ±1/2° SCALE NONE DO NOT SCALE DRAWG. SUPPLIES BY SUPPLIED BY		NO. ZONE RECORD OF CHANGES BY CHKD APPD DATE	303XXXX 1 OF 2