



Certificate of Compliance

Certificate: 1869869 (LR 65677)

Master Contract: 175984

Project: 1869869

Date Issued: 2007/03/07

Issued to: **Mettler-Toledo (Albstadt) GmbH**
Unter dem Malesfelsen 34
Albstadt-Ebingen, 72458
Germany
Attention: Mr. Achim Conzelmann

The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'



Issued by: Eknath Desai; C.E.T.

Authorized by: Renzo Pupulin, C.E.T., Product Group Manager

PRODUCTS

CLASS 8721 84 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - Certified to US Standards

CLASS 8721 04 - LABORATORY EQUIPMENT - Electrical

Scale terminal, Model ID30-Elobox, ID-30 Elobox-2, HMI Box, ID-30 HMI Box-2, ID-30 HMI Box-3 and ID30-HMI-Box-6, non detachable cord connected, rated 100V-240VAC, 50-60Hz, 550mA-250mA

The above model is Equipment Class I, Pollution Degree 2, Installation Category II, portable device and continuous operation.

APPLICABLE REQUIREMENTS

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognized to perform certification to U.S. Standards.



Certificate: 1869869 (LR 65677)

Master Contract: 175984

Project: 1869869

Date Issued: 2007/03/07

CAN / CSA C22.2 No. 61010-1-04 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements

UL 61010A - 1 (Second Edition) - Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use - Part 1: General Requirements

CONDITIONS OF ACCEPTABILITY

The equipment is supplied with an approved power supply cord set or power supply cord with plug that is acceptable to the authorities in the country where the equipment is to be used.