

Adresse Im Hackacker 15 (Industrie Nord), 8902
 Briefadresse Postfach, CH-8902 Urdorf
 Telefon +41-44-729 62 11
 Direktfax +41-44-729 66 26 (Beschaffung)
 MWST-Nr. 226 895

www.mtpro.com

To whom it may concern

Material Specification

Materialbescheinigung

We/ Wir**Mettler-Toledo GmbH, Process Analytics**

Im Hackacker 15
 8902 Urdorf
 Switzerland

hereby confirm that we have attained the Quality Assurance Certificate ISO 9001 and the Environmental Certificate ISO 14001 and that the said product fulfils the following material specification:

bestätigen, dass wir im Besitz des Qualitätszertifikates nach ISO 9001 und des Umweltzertifikates nach ISO 14001 sind und dass die genannten Produkte unserer Firma in Übereinstimmung mit folgenden Materialspezifikationen hergestellt wurden:

Description
Beschreibung**Certificate of compliance**
WerksbescheinigungDescription of Goods
Produktbeschreibung**52403173 Seal Set Kalrez FDA 7XX**
52403173 Ersatzteil-Set Kalrez FDA 7XX

Containing:

| | |
|--------------------|---|
| 52750104 | O-Ring 17.86x2.62 Kalrez FDA¹ |
| 203041016IG | O-Ring 20.29x2.62 Kalrez FDA¹ |
| 203041034IG | O-Ring 10.77x2.62 Kalrez FDA¹ |
| 203041042IG | O-Ring 18.77x1.78 Kalrez FDA¹ |
| 52750150 | O-Ring 53.70x1.78 Vi FDA² |
| 203021131IG | O-Ring 25.12x1.78 Vi FDA² |
| 52750146 | Scraper Ring/ Abstreifring 20.0x2.9x1.1 PTFE FDA³ |

Specifications
Spezifikationen

| | |
|-------------------|---|
| - Material | ¹ Kalrez FDA USP Class VI Compound 6230 according to 21 CFR 177.2600 and according to USP Class VI |
| | ²Viton FDA according to 21 CFR 177.2600 |
| | ³PTFE FDA according to 21 CFR 177.1550 |

Mettler-Toledo GmbH, Process Analytics


Waldemar Rauch
 General Manager PO Urdorf

Place and Date of issue/
Ausstellungsart/- Datum**Urdorf, March 23rd, 2006****METTLER TOLEDO**