



## SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

METTLER-TOLEDO RAININ, LLC  
150 Wells Ave  
Newton, MA 02459  
Henri Chahine Phone: 510 564 1705  
Email: [henri.chahine@rainin.com](mailto:henri.chahine@rainin.com)

### CALIBRATION

Valid To: March 31, 2024

Certificate Number: 2161.04

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

#### I. Mechanical

Parameter/Equipment	Range	CMC <sup>2</sup> (±)	Comments
POVA (Piston Operated Volumetric Apparatus) –  Piston Pipettes, Piston Burettes, Dilutors and Dispensers	(> 0.1 to ≤ 2) µL (> 2 to ≤ 10) µL (> 10 to ≤ 20) µL (> 20 to ≤ 100) µL (> 100 to ≤ 200) µL (> 200 to ≤ 500) µL (> 500 to ≤ 1000) µL (> 1 to ≤ 2) mL (> 2 to ≤ 5) mL (> 5 to ≤ 10) mL (> 10 to ≤ 20) mL (> 20 to ≤ 50) mL (> 50 to ≤ 100) mL (> 100 to ≤ 200) mL	0.016 µL 0.022 µL 0.044 µL 0.27 µL 0.38 µL 1.6 µL 2.2 µL 3.5 µL 7.1 µL 13 µL 22 µL 30 µL 50 µL 65 µL	Gravimetric method per ISO-8655-6

<sup>1</sup> This dedicated field laboratory offers commercial calibration service and field calibration service.

<sup>2</sup> Calibration and Measurement Capability Uncertainty (CMC) is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards or nearly ideal measuring equipment. CMCs represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.



# Accredited Laboratory

A2LA has accredited

**METTLER TOLEDO-RAININ, LLC**

Newton, MA

for technical competence in the field of

**Calibration**

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017 General requirements for the competence of testing and calibration laboratories. This laboratory also meets R205 – Specific Requirements: Calibration Laboratory Accreditation Program. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system  
(refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



Presented this 14<sup>th</sup> day of March 2022.

A blue ink signature of a person's name, which is the Vice President of Accreditation Services.

Vice President, Accreditation Services  
For the Accreditation Council  
Certificate Number 2161.04  
Valid to March 31, 2024

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.