

MODULE COMBINATIONS				DIMENSIONAL CHART					
SCALE SIZE	FIRST REF "D" DIM	MIDDLE REF "E" DIM	TERMINAL REF "F" DIM	"A" FDN LENGTH	"B" OVERALL LENGTH	"C"	"D"	"E"	"F"
40'	15'	10'	15'	40'-1 1/2"	40'-0 1/2"	39'-5 1/2"	14'-10 1/2"	9'-8 1/2"	14'-10 1/2"
45'	15'	15'	15'	45'-3 1/2"	45'-2 1/2"	44'-7 1/2"	14'-10 1/2"	14'-10 1/2"	14'-10 1/2"
47.5'	15'	17.5'	15'	47'-9 1/2"	47'-8 1/2"	47'-1 1/2"	14'-10 1/2"	17'-4 1/2"	14'-10 1/2"
50'	17.5'	15'	17.5'	50'-3 1/2"	50'-2 1/2"	49'-7 1/2"	17'-4 1/2"	14'-10 1/2"	17'-4 1/2"
52.5'	17.5'	17.5'	17.5'	52'-9 1/2"	52'-8 1/2"	52'-1 1/2"	17'-4 1/2"	17'-4 1/2"	17'-4 1/2"

NOTES:

- FOUNDATION LENGTH DIMENSION SHOWN IS NOMINAL. ACTUAL TOLERANCE IS -0 +1"
- APPLY LOCTITE #242 TO ALL GAP COVER BOLTS
- APPLY A 1/4" BEAD OF DIELECTRIC COMPOUND IN THE MALE CONNECTOR OF POWER CELL BEFORE FINAL CONNECTION.
- IF DIELECTRIC COMPOUND BECOMES CONTAMINATED, CLEAN CONNECTOR WITH 68004258 CONNECTOR CLEANER. REAPPLY NEW COMPOUND TO THE CONNECTOR.

CONCRETE NOTES:




- USE 4000 PSI CONCRETE AT 28 DAY AGE. AIR ENTRAINMENT SHALL BE 5%-7%. MAXIMUM AGGREGATE SIZE IS 3/4". MAXIMUM SLUMP AS PLACED SHALL BE 4".
- CONCRETE SHALL BE DIRECT CHUTE PLACED AND THOROUGHLY CONSOLIDATED USING A SPUD TYPE VIBRATOR.
- USE OF CALCIUM CHLORIDE ADMIXTURE IS NOT PERMITTED.
- AFTER FINISHING, A STYRENE BUTADIENE TYPE (30% SOLIDS MIN) CURING COMPOUND SHALL BE APPLIED. EXAMPLES: SUPER FLOOR COAT OR SUPER PLOICURE BY EUCLID; MASTER SEAL 66 BY MASTER BUILDERS.
- REFER TO METTLER TOLEDO SPECIFICATION 61807824 FOR FURTHER INFORMATION.
- ESTIMATED CONCRETE PER MODULE IN CUBIC YARDS. (REF ONLY)
15' X 11' = 3.8 17'6 X 11' = 4.5

NOTES FOR HAZARDOUS AREA WIRING:

- THE HAZARDOUS AREA WIRING AND LIGHTNING PROTECTION KIT IS SPECIFIC FOR THE SUPPLY VOLTAGE AND THE INDICATOR USED. IT MUST BE INCLUDE FOR INTRINSICALLY SAFE COMPLIANCE IN HAZARDOUS LOCATIONS.

IMPORTANT:

THE INTERIOR SURFACE OF ALL LOAD CELL RECEIVERS (BOTH TOP AND BOTTOM) MUST BE GREASED WITH 68004326 DURING CELL INSTALLATION.



16	1	68004258	MTX CLEANER, CONNECTOR
15	32	68004325	ANCHOR BOLT EXPANSION 3/4-10 x 5.5
14	3	30465417	CAPPLUGS KOP 7560CD
13	1	30236780	INSTALL KOP VTX2X1 FLUIDS 3-4 MODS
12	1	30465416	COUPLING KOP 7560
11	1	30465415	CHECKING KOP 7560
10	1	61086717 30351769	KOP PDX WIRING 8LC D2J5 KOP GDD WIRING VTX2X1/7560 3-MOD
9	1	TERM	STANDARD TERM MODULE
8	1	MID	STANDARD MIDDLE MODULE
7	1	FIRST	STANDARD FIRST MODULE
6	8	42904891 72236271	LOAD CELL, PDX, 50+ LOAD CELL, GDD, 30+
5	4	30465438 30465439	LOAD CELL ASSEMBLY 7560 KOP PDX 2LC LOAD CELL ASSEMBLY 7560 KOP GDD 2LC
4	2	68004241	BUMPER BASEPLATE RIGHT
3	2	68004240	BUMPER BASEPLATE LEFT
2	4	68004239	BASEPLATE
1	1	61071325	DATA PLATE

ITEM	QTY	PART NUMBER	DESCRIPTION
MAJOR BILL OF MATERIAL			
REV CHANGE BY DATE SCALE .25 DATE 02/24/18 METTLER TOLEDO			
DRN NLD APPD KRS TITLE 7560CD GLO PDX/GDD 3-MOD			
UNLESS OTHERWISE SPECIFIED, ALL DIMENSIONS ARE IN INCHES, AND DIMENSIONAL TOLERANCES ARE:			
FRACTIONAL DECIMAL ANGULAR THIS PRINT IS FURNISHED WITH THE UNDERSTANDING THAT THE ESSENCE THEREOF WILL NOT BE REPRODUCED IN WHOLE OR IN PART WITHOUT WRITTEN AUTHORIZATION OF METTLER-TOLEDO, LLC. ALL DESIGNS ARE THE PROPERTY OF METTLER-TOLEDO, LLC, AND WILL BE PROTECTED BY PATENTS.			
1/32 .XX ±.02 ±.5° 61809014 REV			
1/32 .XXX ±.005 ±.5°			