

Certificate for MVQ 75.24-01

Werkstoff: MVQ
Farbe: transparent
Bezeichnung: MVQ 75.24-01

Date: 13/5/04

For the attention of: Mettler-Toledo GmbH (Harald Hinzert)

We Hereby Certify

That the following grade has been formulated in accordance with United States Food and Drug Administration (FDA) Code of Federal Regulations Title 21 (CFR21) Section 177.2600.

The material is compliant with the following requirement:

- USP Biological Test classification VI.
- FDA 177.2600
- BGV XV
- WRAS (WRC) Certificate Number: 0005509.

| <u>Grade</u> | <u>Material Description</u> |
|-----------------|---|
| Platinum Cured, | Translucent Silicone Rubber (VMQ) 75 IRHD |

Angst+Pfister AG
Thurgauerstrasse 66
CH-8052 Zürich
Telefon 01/306 61 11
Telefax 01/302 18 71

Angst+Pfister SA
52, rte du Bois-des-Frères
CH-1219 Genève-Le-Lignon
Telefon 022/796 42 11
Telefax 022/796 07 78

Logistikcenter
Hardholstrasse 31
CH-8424 Embrach
Telefon 01/866 61 11
Telefax 01/866 65 22

Angst+Pfister SpA
Viale Teodorico 25
I-20149 Milano
Telefon 0039/23 10 61
Telefax 0039/233 10 31 48

Angst+Pfister GmbH
Siemenstrasse 2
D-64546 Mörfelden/Ffm
Telefon 0049/610 59 25 40
Telefax 0049/610 52 54 69

Angst+Pfister Ges.m.b.H.
Leopoldauerstrasse 143
A-1210 Wien
Telefon 0043/125 84 60 10
Telefax 0043/125 84 60 198

Angst+Pfister SA
Ch. de Goulle Cédex
F-95950 Raissy
Telefon 0033/148 63 20 80
Telefax 0033/148 63 26 90
