

Address Im Hackacker 15, CH-8902 Urdorf, Switzerland
 Mail address P.O. Box, CH-8902 Urdorf, Switzerland
 Phone +41-44-729 62 11
 Fax +41-44-729 66 36
 Bank Credit Suisse, 8070 Zurich, Clearing 4835
 Account No. 370501-21-90 CHF/IBAN CH71 0483 5037 0501 2109 0

To whom it may concern

Material Specification EN 10204-2.1
Materialbescheinigung EN 10204-2.1

We/ Wir

Mettler-Toledo AG, Process Analytics

Im Hackacker 15
 8902 Urdorf
 Switzerland

hereby confirm that we have attained the Quality Assurance Certificate ISO 9001 and the Environmental Certificate ISO 14001 and that the said product fulfils the following material specification:

bestätigen, dass wir im Besitz des Qualitätszertifikates nach ISO 9001 und des Umweltzertifikates nach ISO 14001 sind und dass die genannten Produkte unserer Firma in Übereinstimmung mit folgenden Materialspezifikationen hergestellt wurden:

Description

Certificate of compliance with the order EN 10204-2.1

Beschreibung

Werksbescheinigung EN 10204-2.1

Description of Goods

InPro7002/ TC –VP 3.1B and InPro7002/** DN** types**

Produktbeschreibung


Specifications

Medium wetted parts / mediumberührte Teile:

Spezifikationen

Metallic and/or polymer parts	
shaft	Stainless steel AISI 316L 3.1B electropolished: Ra <0.2 µm (<8 µinch)
electrodes	Stainless steel AISI 316L 3.1B electropolished: Ra <0.2 µm (<8 µinch) except active measuring areas
Insulator between electrodes	PEEK FDA according to 21 CFR 177.2415/ sanitary standard 20-19 and directive 93/9/EEC
O-Rings	Viton FDA according to 21 CFR 177.2600

Mettler-Toledo AG, Process Analytics



Waldemar Rauch
 General Manager PO Urdorf

Date/Datum

November 24, 2006