Mettler-Toledo GmbH
Process Analytics

Adresse | Im Hackacker 15 (Industrie Nord), CH-8902 Urdorf, Schweiz
Briefadresse | Postfach, CH-8902 Urdorf
Telefon | 01-736 22 11
Telefax | 01-736 26 36
Internet | www.mt.com
Bank | Credit Suisse First Boston, Zürich (Acc. 0835-370501-21-90)

To whom it may concern

Material Specification EN 10204-2.1
Materialbescheinigung EN 10204-2.1

Wer/ Wir

Mettler-Toledo GmbH, Process
Im Hackacker 15
8902 Urdorf
Switzerland
hereby confirm that we have attained the Quality Assurance Certificate ISO 9001 and the Environmental Certificate ISO 14001 and that the said product fulfils the following material specification:
bestätigen, dass wir im Besitz des Qualitätszertifikates nach ISO 9001und des Umwelt-zertifikates nach ISO 14001 sind und dass die genannten Produkte unserer Firma in Übereinstimmung mit folgenden Materialspezifikationen hergestellt wurden:

Description
Beschreibung

Certificate of compliance with the order EN 10204-2.1
Werksbescheinigung EN 10204-2.1

Description of Goods
Produktbeschreibung

InFlow76X

Specifications
Spezifikationen

Medium wetted parts/ mediumberührte Teile

	InFlow 761
Metallic parts:	Stainless steel DIN 1.4404
	InFlow 762
Polymer parts:	PVDF
O-rings:	Only for type with housing connection Ingold DN25: Viton FDA according to 21 CFR 177.2600

Mettler-Toledo GmbH, Process Analytics

Waldemar Rauch
General Manager PO Urdorf

Date

Juni 13, 2005

METTLER TOLEDO

Sitz der Gesellschaft Mettler-Toledo GmbH, Im Langacher, CH-8606 Greifensee