

Adresse Im Hackacker 15 (Industrie Nord), 8902
 Briefadresse Postfach, CH-8902 Urdorf
 Telefon +41-44-729 62 11
 Direktfax +41-44-729 66 26 (Beschaffung)
 MWST-Nr. 226 895

www.mtpro.com

To whom it may concern

Material Specification EN 10204-2.1

Materialbescheinigung EN 10204-2.1

We/ Wir**Mettler-Toledo GmbH, Process**

Im Hackacker 15
 8902 Urdorf
 Switzerland

hereby confirm that we have attained the Quality Assurance Certificate ISO 9001 and the Environmental Certificate ISO 14001 and that the said product fulfils the following material specification:

bestätigen, dass wir im Besitz des Qualitätszertifikates nach ISO 9001 und des Umweltzertifikates nach ISO 14001 sind und dass die genannten Produkte unserer Firma in Übereinstimmung mit folgenden Materialspezifikationen hergestellt wurden:

Description**Certificate of compliance with the order EN 10204-2.1**Beschreibung**Werksbescheinigung EN 10204-2.1**Description of GoodsProduktbeschreibung**52403583 Diary Adapter/ Milchrohradapter for InPro7250****52403584 Aseptic Adapter/ Aseptik-Adapter for InPro7250**SpecificationsSpezifikationen**Medium wetted parts/ mediumberührte Teile**

	Diary Adapter Milchrohradapter	Aseptic Adapter Aseptik-Adapte
Metallic parts: Metallische Teile:	Stainless steel DIN 1.4435 Rostfreier Stahl DIN 1.4435	Stainless steel DIN 1.4435 Rostfreier Stahl DIN 1.4435
O-rings:	EPDM FDA Compound EPDM 70.10-02 according to 21 CFR 177.2600	EPDM FDA Compound EPDM 70.10-02 according to 21 CFR 177.2600

Mettler-Toledo GmbH, Process Analytics


Waldemar Rauch
 General Manager PO Urdorf

Date**March 23rd, 2006**

MS_21_Adapter_InPro7250_52403583_52403584_March06_int 1 (1)

METTLER TOLEDO