



Member of the FM Global Group

FM Approvals
1151 Boston Providence Turnpike
P.O. Box 9102 Norwood, MA 02062 USA
T: 781 762 4300 F: 781-762-9375 www.fmapprovals.com

CERTIFICATE OF COMPLIANCE

HAZARDOUS LOCATION ELECTRICAL EQUIPMENT PER CANADIAN REQUIREMENTS

This certificate is issued for the following equipment:

SLB615D. Digital Weighing Load Cell.

NI / I, II / 2 / ABCDFG / T6 Ta = 55°C — 30095704; NIFW

DIP / III / 2 / T6 Ta = 55°C

Nonincendive Field Wiring Parameters:

VMax = 28V, IMax = 50 mA, Ci = 0.3 nF, Li = 0

Equipment Ratings:

Nonincendive with NIFW parameters for Class I, II, Division 2, Groups A, B, C, D, F, and G; Dust Ignition-proof for Class III, Division 2; indoor and outdoor Hazardous Locations

FM Approved for:

Mettler Toledo
Changzhou, China



Member of the FM Global Group

This certifies that the equipment described has been found to comply with the following Approval Standards and other documents:

CSA C22.2 No. 213	1987 (R2008)
C22.2 No. 1010.1	2004
C22.2 No. 25	1966 (R2004)
CSA C22.1	2009

Original Project ID: 3050641

Approval Granted: May 6, 2014

Subsequent Revision Reports / Date Approval Amended

Report Number	Date	Report Number	Date
---------------	------	---------------	------

FM Approvals LLC



J.E. Marquedant
Manager, Electrical Systems

6 May 2014

Date

To verify the availability of the Approved product, please refer to www.approvalguide.com